



RED Declaration of Conformity 2014/53/EU

We, Unitron Hearing a division of National Hearing Services Inc., located at 20 Beasley Dr., Kitchener, ON, Canada, N2E 1Y6, declare under our sole responsibility that the following radio equipment (wireless hearing aids):


| Unitron | Vista | AudioNova |
|----------------------------|----------------------|-------------------------|
| DX Moxi Move R 3 | Vista 350 DX R Li | AudioNova DX 30 R Li |
| DX Moxi Move R 5 | Vista 550 DX R Li | AudioNova DX 50 R Li |
| DX Moxi Move R 7 | Vista 750 DX R Li | AudioNova DX 70 R Li |
| DX Moxi Move R 9 | Vista 950 DX R Li | AudioNova DX 90 R Li |
| DX Moxi Move R FLEX:TRIAL | Vista DX R Li TRIAL | AudioNova DX R Li TRIAL |
| DX Moxi Move R FLEX:SELECT | Vista SELECT DX R Li | |

Device Classification: IIa, Notified Body: Presafe Denmark A/S, CE Certificate Number: DGM-046
 Authorized Representative: Sonova AG, Laubisrütistrasse 28, CH-8712, Stäfa, Switzerland


are in conformity with the requirements of the Radio Equipment Directive 2014/53/EU, based upon compliance of the products with the following standards:

| Requirement | Standard |
|-----------------------------------|--------------------------------|
| Health & Safety (Art. 3(1)(a)) | EN/IEC 60601-1:2006 |
| | EN/IEC 62479:2010 |
| EMC (Art. 3(1)(b)) | EN 60601-1-2:2015 |
| | ETSI EN 301 489-1 V2.2.0 |
| | ETSI EN 301 489-3 V2.1.1 |
| | EN 301 489-17 V3.2.0 (2017-03) |
| RADIO SPECTRUM (Art. 3(2)) | ETSI EN 300 328 V2.1.1 |

Placed on the market (CE marked): [April 1 2020](#)



Ara Talaslian
Vice President, R&D



Craig Horsman
Senior Manager, Quality Assurance

Kitchener, ON *MARCH 12, 2020*